ISSN: 2356-7767

Validation of HPLC methods for detection of selenium and Nano-selenium in chicken tissues

Mai, A. Fadel^{*}; Fadwa, F. Mahmoud^{**} and Essam, S. Soliman^{***}

 *Reference Laboratory for Veterinary Quality Control on Poultry Production (RLQP), Animal Health Research Institute (AHRI), Dokki, Giza, Egypt
 **Reference Laboratory for Veterinary Quality Control on Poultry Production (RLQP), Ismailia, Egypt, 41513

*** Department of Animal Hygiene, Zoonosis, and Animal Behavior, Faculty of Veterinary Medicine, Suez Canal University, Ismailia, Egypt, 41522

Received in 12 /2/2019 Accepted in 19/3/2019

Abstract

As a part of an ongoing study to develop and validate an accurate, sensitive and reproducible reverse phase high performance liquid chromatographic (RP-HPLC) method with a UV detector for the quantization of selenium (Se) and nano-selenium (N-Se) in chicken tissues (muscle and liver) following oral supplementation of Se and N-Se at a dose 100mg/L. Pre-column derivatization and reversed-phase chromatography provides separation of nano-selenium compound and eluted from a C8 column (4.6 mm i.d.,150 mm, 5 μ m particle size, pore size100 Å) and the mobile phase was methanol (85%):10mM citric acid monohydrate (15%) with adjusted pH of citrate buffer at 2.5.The samples were extracted for selenium detection by perchloric and nitric acids. Separation of selenium compound was carried out onC18 column (4.6 mm, i.d., 250 mm, 5 μ m) with a mobile phase 100% methanol. A UV detector set at 332 nm and 280nm were used to monitor the effluent (nanoselenium and selenium), respectively. The limit of detection and quantification was found to be 0.014032 µg/ml,0.046775µg/ml and 0.008657µg/ml, 0.026234µg/ml for Se and N- Se, respectively. Linearity for N-Se was in a range of 0.01 to 2.5 µg/ml and for Se was 0.005 to 2.5 µg/ml. The proposed methods were highly sensitive, accurate and precise and could be used for routine analysis of N-Se and Se in chicken tissues.

Keywords: RP, HPLC, Selenium, Nano Selenium, Chicken tissues.

Introduction

Selenium is a trace element essential in animal nutrition and exerts multiple actions related to animal production, fertility, and disease prevention (Mervyn,1985). Selenium is an integral part of the enzyme glutathione eperoxidase, which serves as an antioxidant enzyme that helps to control levels of hydrogen peroxide and lipid peroxides that are produced during normal metabolic activity (**Rotruck** *et al.*, **1973).** In addition, dietary selenium is essential for the activity of virtually all arms of the immune system (**Suraiand Dvorska**, **2002**). Selenium, like all biologically essential trace elements, can be toxic when provided at levels in excess of the biological requirement. Trace elements typically demonstrate what has been referred to in toxicology as a Ushaped response curve (Hayes, 2008), which describes the negative effects of selenium deficiency and the negative effects of selenium at excessive dietary inclusion levels. Selenium has long been known to be toxic and there are concerns about its effect on animals and animal products. To ensure feed safety, maximum levels for selenium in complete feeds have been set at 0.5 mg/ kg in the European Union (2004) and China (Ministry of Agriculture, 2010) and 2.0 mg/ kg for the United States (AAFCO, 2011).

Nano-materials exhibit novel properties, such as great specific surface area, high surface activity, a lot of surface active centers and high catalytic efficiency (Gao and Hiroshi, 2005). Due to the advantage of size effect and high surface reactivity, nanoparticle has been already used in pharmaceutical applications to increase the bioavailability of drugs and targeting therapeutic agents to particular organs (Florance et al., 1995; Davda and Labhasetwar, 2002). Nano red elemental selenium (nano- Se) of size 5-100 nm was synthesized by (Zhang et al. 2001) and observed that nano -Se had a similar bioavailability in rat and much less acute toxicity in mice compared with selenite. It has been reported that nanoshowed new characteristics particle of transport and uptake and exhibited higher absorption efficiencies (Davda and Labhasetwar, 2002; Chithrani and Chan, 2007; Zha et al., 2008; Liao et al., 2010). Advantages of ultrasound-assisted extraction of trace elements from a variety of matrices using both bath and probe ultrasonic processors technique have been widely recognized for efficient separation of metals and metalloids from a variety of solid matrices, hence avoiding intensive and destructive sample treatments such as dry ashing or acid digestion procedures. Ultrasonic probes provide some advantages in comparison with ultrasonic baths, namely, short extraction times, extraction efficiency approaching 100% and feasibility of using low acidic conditions for extraction as a result of the enhanced ultrasonic power delivered by the former (i.e., high intensity sonication), which can be directly focused onto the sample (Bendicho & Lavilla, 2000). The aim of this study is to develop and validate new methods for extraction and quantification of selenium and its Nano form by VIS-UV RP-HPLC in chicken tissues. These methods enabled the quantification of selenium and its nano form residues and their withdrawal time to be safe for human consumption to avoid selenium toxicity.

Materials and methods a-Instrumentation:

The high-performance liquid chromatography system consisted of Agilent Series 1200 qua-

ternary pump, Series 1200 auto sampler, Series 1200 UV Vis detector and HPLC 2D Chemstation Software (Hewlett-Packard, Les Ulis, Germany).

The separations were performed on reversedphase (RP) column C18 (4.6 mm, i.d., 250 mm, 5 μ m), (Kromasil Co.) and (RP) column C8 (5 μ m particle size, pore size 100 Å, L × I.D. 150 mm × 4.6 mm

Cooling centrifuge, ultrasonic bath (Buhler, Germany), vortex mixer (Inc., N.Y., USA) and shaker (Vibra).

b-Reagents, standards and mobile phase:

All commercial chemicals were HPLC grade. Selenium dioxide powder (SeO2) 1000 mg/L, Titrisol, 9915) from Merck. Selenium Nan powder with a particle size (50nm) (NS613010 -171) from Nanoshel.

Methanol, Hexane and Toluene were of HPLC grade, while other chemicals and reagents analytical grade were obtained from BDH Laboratories Supplies (BDH Chemical Ltd., Poole, U.K.).

Deionized, Milli Q water (Millipore, Bedford, MA, USA) was used to prepare the mobile phase and diluent solutions.

c-Chromatographic separation:

Isocratic mobile phase for selenium separation was 100% MeoH (HPLCgrade) and for Nanoselenium separation consisted ofMethanol (85%),10mM citric acid monohydrate (15%) with adjusted pH of citrate buffer at 2.5.

d-Standard preparation:

A stock standard solution of 1 mg/ml was prepared by dissolving 10mg of selenium standard in 10 ml of diluted nitric acid (2%) and the same amount for nano selenium standard was dissolved in 10 ml de-ionized water.

e-Assay procedure

d-1-Extraction method for nano-selenium:

The homogenized tissue sample(1 g) was weighed in a 50 ml propylene tube and add 4

ml of 10% Trichloroacetic acid (TCA), 1ml of 30%Nitric acid and 20µl concentrated sulfuric acid (H₂SO₄)then vortex for 30 seconds. The samples were heated in water bath at 90° C for 20 minutes, then 0.1 g of sodium sulfite was added, the samples were ultra-sonicated for 30 minutes the ncentrifugedat 2000 rpm for 10 minutes. The supernatant was filtrated through an acid washed filter paper (TCA 10%). The filtrate pH was adjusted to 1.0 by adding 0.1NHCland followed by the derivatization steps (Arain et al., 1999). DAB (2,3diaminobenzene) reagent (2ml with 1% 0.1N HCL) was added to the sample filtrate, and the tubes were tightly capped. The contents were mixed well, and the tubes placed in a dark at 30° C in water bath for 40 min for synthesis of 2,1,3 benzoselenadiazole (BSD). After that, 1ml of toluene was added and vortex for 30 seconds and centrifuged at 2000rpm for 5 minutes. The organic phase was collected and evaporated at 40° C and the residues eluted by 200 μ l of methanol: water (50:50 v/v). The HPLC sample injection volume was 10µl at flow rate 1ml/min. with UV detection at 332 nm (Arain et al., 1999) and mobile phase wasmethanol:10mM citric acid monohydrate (85:15) with adjusted pH of citrate buffer at 2.5.

d-2- Extraction method for selenium:

One gram of homogenized tissue sample was weighed in 50 ml falcon tube. Onemilliliterof perchloricacid, 1ml of Nitric acid 65% and 50µl concentrated sulfuric acid were added then vortex for 60 seconds. The samples were heated on a hotplate at 100° C till the white fumes of perchloric acid were evolved. The sample volume was maintained approximately at 1 ml to prevent the volatilization of selenium. Two milliliter of hexane was added and vortex for 30 seconds and centrifuged at 2000rpm for 5 minutes. The organic phase was collected and evaporated at 40° C.The dried residues were dissolved by 200 µl of a mixture of methanol:2 propanol(1:1). The HPLC sample injection volume was 50µl at flow rate 1ml/min. with UV detection at 280 nm and mobile phase was methanol (HPLC grade).

Method validation

It is the procedure by which the performance characteristics of the method meet the requirements for the intended analytical application. The proposed method was validated according to the International Conference on Harmonization (ICH) guidelines. ICH, Q2 (R1), Harmonised tripartite guideline, Validation of analytical procedures: text and methodology International Conference Harmonization ICH, Geneva, Nov 2005.

System suitability

It is an integral test of many analytical procedures. In chromatography, there are some parameters which are column plates (theoretical plates) for judging column efficiency, peak asymmetry factor (As), tailing factor (T_f) and resolution (Rs). Rs is a measure of the separation quality or in other words; is specified to ensure that closely eluting compounds are resolved from each other and according to the united states pharmacopeia (USP, 2017).

Precision

2.A.The intraday precision

Precision is the closeness of agreement among a set of results. The intraday precision (Repeatability) of the assay were measured by analyzing six replicate of spiked tissue sample with selenium ($0.02\mu g/mL$) and nano-selenium ($0.08\mu g/mL$) on the same day. The percent of relative standard deviation (% RSD) for peak responses was calculated according to (ICH,2005).

2. B. Intermediate precision:

Intermediate precision (within-laboratory variation) determined by estimation of six replicates of spiked tissue samples with selenium (0.02µg/mL) and nano-selenium (0.08µg/mL) daily for 6 times over a period of one week (interday precision). The %RSD for peak responses was calculated according to (ICH, 2005)

Linearity and range

Linearity was performed by preparing an eight different concentrations of drug standard at squared correlation coefficient of 0.99 (r^2) according to ICH. Calibration lines of peak area ratios (peak area analyte/peak area internal standard) versus concentration were determined by single level calibration curve (linear regression equation, Y = A X +B), where: Y =

peak area ratio, A = slope, X = concentration and B = intercept.

Specificity:

Solutions of standard and spiked tissue samples were prepared as per the test method and injected into the chromatographic system, the chromatograms were recorded.

Accuracy and Recovery:

The accuracy of an analytical method is the closeness of test results (theoretical value) obtained by method to the assay value. Accuracy must be established across the specified range of the analytical procedure. Accuracy was determined over the range of 50%, 100% and 150% of the sample concentration according to **(ICH, 2005)**.

The accuracy was then calculated as the percentage of analyte recovered by the assay.

% Recovery = [Theoretical value / Assay value] × 100

Accepted criteria of recovery ranged from 85-110%

Robustness:

It was determined by detecting how a method stands up to slight variations in normal operating conditions. The chromatographic conditions change in selenium method were studied. The effect of mobile phase composition was assessed at 93% methanol instead of 100% methanol

The determination of nano-selenium method robustness, small deliberate changes in the chromatographic conditions, such as the effect of mobile phase composition was assessed at methanol: **citrate buffer (87:13 v:v)** and methanol: **citrate buffer (83:17 v:v)** instead of methanol: **citrate buffer (85:15 v:v)** .The effect of pH of citrate buffer of mobile phase at 2.3 and 2.7 instead of 2.5. The effect of column (C18 instead of C8) was studied. The %RSD of robustness testing under these conditions was calculated in all cases.

Limits of detection (LOD)

It is considered to be the quantity yielding a detector response which gives signal to noise ratio 3:1 according to (ICH,2005)

Limits of quantification (LOQ)

It is the lowest amount that can be analyzed within acceptable precision and accuracy which give signal to noise ratio 10 : 1 according to (ICH, 2005)

Results and Discussion System suitability

Column efficiency for selenium peak was 7500 plates and for nano selenium was 9700 plates as Theoretical plates (N) according to USP is = \geq 2000. Peak asymmetry factor (As) was 1.158 for selenium and for nano selenium was 0.95 which considered fronting as As was less than 1. Tailing factor as USP coefficient (T = \leq 2.0)of the peak symmetry was 1.03 and 0.99 for selenium and nano selenium, respectively.

Precision

Precision of the developed method was determined based on inter and intra-day precisions. Results are presented in Table (1). The method was found to be precise since the RSD values for both inter-day and intra-day precision was below 1.0

2.a. Intraday Precision

Selenium			Nano-selenium		
Sr. no.	Concentration (µg/ml)	Peak area (In the same day)	Concentration (µg/ml)	Peak area (In the same day)	
1	0.02	40.84	0.08	38.9	
2	0.02	40.77	0.08	38.029	
3	0.02	40.4	0.08	38.3111	
4	0.02	40.59	0.08	38.205	
5	0.02	40.7	0.08	38.57778	
6	0.02	40.34	0.08	38.6127	
Mean		40.64		38.43926	
SD		0.2		0.316596	
RSD%		0.5		0.823628	

2.b. Inter day precision value:

Selenium			Nano-selenium		
Sr. no.	Concentration (µg/ml)	Peak area (In six days)	Concentration (µg/ml)	Peak area (In six days)	
1	0.02	40.014	0.08	38.3098	
2	0.02	39.887	0.08	38.72	
3	0.02	39.942	0.08	38.804	
4	0.02	39.975	0.08	38.88	
5	0.02	40.005	0.08	37.943	
6	0.02	40.42	0.08	38.775	
Mean		40.041		38.57197	
SD		0.192		0.36775	
RSD%		0.478		0.95341	

linearity and range

The matrix-matched calibration standards were prepared by spiking the blank material with the same spiking solution as for samples for validation with variable volumes. Standard curves were constructed by spiking of eight blank muscle and liver tissue broiler samples with various volumes of selenium and nanoselenium stock solution to yield a concentration range of 0.005, 0.02, 0.04, 0.08, 1.0, 0.25, 0.5, 1 and 2.5 μ g/ml and 0.01, 0.02, 0.04, 0.08, 0.1, 0.25, 0.5, 1 and 2.5 μ g/ml, respectively.

The Standard curves were shown linearity from 0.005 to 2.5μ g/ml for selenium standard in solvent fig. (1-b), muscle tissue samples fig. (1-d) and in liver tissue fig. (1-f). Meanwhile, The linearity from 0.01 to 2.5 μ g/ml for nano-selenium standard in solvent fig (1.a), blank muscle samples figure (1-c) and in liver tissues figure (1-e).

The linearity of peak area responses versus concentrations was demonstrated by linear least square regression analysis. The linear regression equations and correlation coefficient (r^2) were illustrated in **figure1 (a, b, c, d, e & f)**.

Linearity values and peaks area were shown in **table (3).** Analytical performance parameters were shown in **table (4).**

Selenium				Nano-selenium			
RT	Level	concentration (µg/ml)	Peak area	RT	Level	Conc.(µg/ml)	Peak area
3.765	1	0.005	9.8	7.073	1	0.01	2.741
	2	0.02	40.899		2	0.02	8.1
	3	0.04	80.5		3	0.04	18.05
	4	0.08	161.2		4	0.08	39.9
	5	0.1	207.99		5	0.1	50.022
	6	0.25	518.85		6	0.25	128.14
	7	0.5	1001.8		7	0.5	260.42
	8	1	2012.6		8	1	524.84
	9	2.5	5113		9	2.5	1272.5

 Table (3). linearity of Selenium and nano-selenium standard in solvent:

*RT: Retention Time

 Table (4). Analytical performance parameters

Dovomotova	Selenium	Nano-selenium		
Farameters	Value			
Linearity range	0.005-2.5 µg/ml	0.01- 2.5 μg/ml		
Correlation coefficient(r ²)	0.9999	0.9998		
Slope (a)	2042.1	511.04		
Intercept (b)	-0.4839	0.7519		
Regression equation	y=2042.1x-0.4839	Y=511.04x+0.7519		



Figure (1). Matrix-matched calibration plots for nano-selenium (a) and Selenium (b) in solvent , chicken meat (c,d) and Liver tissues (e,f).

Specificity

The chromatograms compared to know that there is no excipient compound interference between peaks of the pure standard and peaks of spiked tissue (muscle and liver) samples. Figure 2.



Figure (2). Chromatogram of spiked blank muscle sample with (a) selenium and (b) nano-selenium (0.1µg/ml)

Accuracy and Recovery

Spiked samples will be prepared at three concentrations of 50, 100 and 150 % which represented by 0.02, 0.04, and 0.08 μ g/ml respec-

Table (5). Recovery & Accuracy studies(a) Selenium

tively. Each spiked sample was prepared in triplicate at each level and injected. The absolute recovery of selenium and nano-selenium were shown in table (5).

Concentration level (μg/ml)	Found Conc. (µg/ml)	Recovery %	Mean ± SD	RSD%	Average % Recovery (Accuracy) ± SD
	0.02004	100.2007	0.02012±		
0.02 (50%)	0.02013	100.6664	0.000074	0.36973	100.601±0.372
	0.02019	100.936			
	0.039657	99.14292	0.02070	0.02105	
0.04 (100%)	0.03951	98.77565	$0.039/9\pm$ 0.00037	0.92103	99 47754+0 92
	0.04021	100.5141	0.00037		JJ.47754±0.72
0.06 (150%)	0.059098	98.49656	0.050415	0.552214	00 00424+0 54792
	0.059392	98.98626	$0.039413\pm$ 0.000328	0.555214	99.02434±0.34782
	0.059754	99.59021	0.000320		

(b) Nano-selenium

Concentration level (µg/ml)	Found Conc. (µg/ml)	Recovery %	Mean±SD	RSD%	Average % Recovery (Accuracy) ± SD
	0.02005	100.26517	0.0198377±		
0.02 (50%)	0.01969	98.45632	0.0001905	0.96	99.1887±0.952
	0.019769	98.84461			
	0.03975	99.37874		0.77731	
0.04 (100%)	0.040017	100.04167	0.039723 ± 0.00 0309		99 30771+0 772
	0.0394011	98.502727	0509		<i>yy.30771</i> ±0.772
	0.0597341	99.5567846			
0.06 (150%)	0.05965831	99.43051241	$0.0598793{\pm}0.0$ 0032	0.53337	99.7988±0.53231
	0.0602451	100.4091219			

Robustness:

The results of robustness indicated that changing the mobile phase composition and changing the pH of citrate buffer of mobile phase had slight effect on the chromatographic behavior of nano-selenium. However, changing the column had no significant effect, also; changing in the mobile phase composition and wavelength had a slight effect. The RSD % of robustness testing under different altered conditions are given in Table (6) and (7), indicating that the current methods were robust.

Chromatographic parameter	Chromatographic parameter Modification		Mean area ± SD	RSD%
Change in Mobile phase composition	10mM citric acid monohy- drate: Methanol 15:85 17:83 13:87	38.44895 37.193573 37.678667	37.77373± 0.63306	1.675938
- H of situate baffer	2.3	37.7513	28 (022 + 0 02497	
of mobile phase	2.5	38.7287	38.0933±0.92487	2.3903
	2.7	39.6		
Change in Column	C8	38.2125	37 95771+0 360332	0 949299
Change in Column	C18	37.70291	57.75771±0.300332	0.979299

Table (6). Results of Robustness study (Nano-selenium concentration, $0.08 \mu g/mL$, n=3)

Table (7). Results of Robustness study (Selenium concentration, 0.02 µg/ mL, n=3)

Chromatographic parameter	Chromatographic Modification		Mean area ± SD	RSD%
Change in Mobile phase composition	Methanol: Water 100% methanol 93:7	39.78451 38.40466	39.09459±0.975701	2.495744
Change in Wave- length(nm)	280 277 283	39.87977 38.72295 39.3635	39.32207± 0.579525	1.47379

Limit of detection and limit of quantification

LOD and LOQ were calculated using the following equations designated by International Conference on Harmonization (ICH) guidelines.

$$LOD = 3.3 \times \sigma/S$$

$$LOQ = 10 \times \sigma/S$$

Where, σ is the standard deviation of intercept S is slope of the calibration curve. So, LOD of Selenium was 0.008657 µg/ ml and LOQ was 0.026234µg/ml. On the other hand; LOD of nano-selenium was 0.014032 µg/ ml and LOQ was 0.046775µg/ml.

Discussion

The development of HPLC methods for detection and quantification of analytes is very important in analytical researches for achieving the quality control. Despite all previous experimental studies for development and validation of analytical methods for determination of selenium and nano-selenium in chicken tissues with different detectors as (Li et al., 1999 and Pedersen & Larsen, 1997) which carried out the separation of selenium compounds by HPLC on an ESA Anion III anion-exchange column and on-line selenium-specific detection was carried out with an inductively coupled plasma mass spectrometer (ICP-MS) or a flame atomic absorption spectrometer (FAAS), these methods achieved the detection of selenium compounds by RP-HPLC through following the validation parameters such as accuracy, precision, robustness, limit of detection (LOD) and limit of quantification (LOQ) according to

ICH guidelines. The developed methods were reliable, simple, sensitive, precise, accurate and economic as using an UV detector. Meanwhile, other developed methodsas (Nakagawa et al., 1989 and Tanaka et al., 1988) which used a fluorometric detector in detection of selenium compounds. The developed method, using simple HPLC grade solvents (methanol and water), had a short retention time and high peak symmetry. The method developed was validated successfully. The selected mobile phase system gave a single sharp peak for each selenium and nano-selenium without interfering peaks. The retention time (RT) of was 3.765 min and 7.073 with a tailing factor 1.03 and 0.99 for selenium and nano selenium, respectively. The method was linear in the concentration range of $0.005 - 2.5 \mu g/ml$ and $0.01 - 2.5 \mu g/ml$ for selenium and nano selenium, respectively, and the calibration curves showed good correlation between concentration and peak area.

Conclusion

In this work, we have presented a new method for the extraction and pre-concentration of selenium and Nano- selenium as standard in solvent and in chicken tissues (muscle and liver). The proposed methods were based on the use of methanol (organic solvent) in selenium validated method to make a compound with selenium which called dimethylsiloxane (SiOcH3).

On the other hand; we used DAB (2,3diaminobenzene) reagent to react with nanoselenium for synthesis of 2,1,3 benzoselenadiazole (BSD). These two compounds allowed the extraction and pre-concentration of selenium and nano-selenium after the derivatization step, and so could be detected by Vis-UV HPLC. The developed and validated Vis-UV HPLC methods were found to be more economical. The result of analysis of formulation and recovery studies obtained by HPLC method were statistically validated and high percentage of recovery studies suggest that the developed methods were free from interference of excipients used in formulation. The Vis-UV HPLC methods were statistically validated in terms of accuracy, precision, linearity and reproducibility. Hence above methods can be employed in laboratories to estimate selenium and nano-selenium in chicken tissues (muscle and liver).

References

- AAFCO (2011). AAFCO Model Guidance Document. Official guidelines suggested for contaminants in individual mineral feed ingredients. Official publication. Association of American Feed Control Officials Inc., Olympia, WA. 304.
- Arain, M.A.; Khuhawar, M.Y. and Bhanger, M.I. (1999). Liquid Chromatographic Determination of Selenium in Vegetables and Tea Leaves As 2, 1, 3-Benzoselenasiazole. J Chem Soc Pak.; 21(2): 137-140.
- Bendicho, C. and Lavilla, I. (2000). Applications of ultrasound-assisted metal extractions. Encylopaedia of Separation Science. London: Academic Press; 2000. pp. 4421–6. General Chapter <621> "Chromatography" in United States Pharmacopeia 40 National Formulary 35 (USP 40–NF 35, United States Pharmacopeial Convention, Rockville, Maryland, (2017); pp. 508–520.
- Chithrani, B.D. and Chan, W.C. (2007). Elucidating the mechanism of cellular uptake and removal of protein-coated gold nanoparticles of different sizes and shapes. Nano Lett.;7(6):1542–50. https://doi.org/10.1021/nl070363yPMID:17465586.
- Davda, J. and Labhasetwar, V. (2002). Characterization of nanoparticle uptake by endothelial cells. Int J Pharm.; 233(1-2):51–9. https://doi.org/10.1016/S0378-5173(01) 00923-1PMID:11897410.
- **European Union. (2004)**. List of the authorized additives in feeding stuffs published in application of Article 9t (b) of Council Directive 70/524/EEC concerning additives in feeding stuffs. Off. J. Eur. Union. C/50:1– 14.
- Florence, A.T.; Hillery, A.M.; Hussain, N. and Jani, P.U. (1995). Nanoparticles as carriers for oral peptide absorption: studies on particle uptake and fate. J Control Release.; 36(1-2): 39–46. https://doi. org/10.

1016/0168-3659(95)00059-Н.

- Gao, X.Y. and Matsui, H. (2005). Peptidebased nanotubes and their applications in bionanotechnology. Adv Mater.; 17(17): 2037–50.https://doi.org/10.1002/ adma.200401849.
- Hayes, D.P. (2008). Adverse effects of nutritional inadequacy and excess: a hormetic model. Am J Clin Nutr.; 88 (2 Suppl.): 578S –81S. https://doi.org/10.1093/ajcn/ 88. 2. 578 SPMID:18689405.
- **ICH (2005)**, Q2 (R1), Harmonized tripartite guideline, Validation of analytical procedures: text and methodology International Conference Harmonization ICH, Geneva.
- Liao, C.D.; Hung, W.L.; Jan, K.C.; Yeh, A.I.; Ho, C.T. and Hwang, L.S. (2011). Nano/sub-microsized lignan glycosides from sesame meal exhibit higher transport and absorption efficiency in Caco-2 cell monolayer. Food Chem.; 119:896–902.
- Li, F.; Goessler, W.; and Irgolic, K.J. (1999). Determination of selenium compounds by HPLC with ICP-MS or FAAS as seleniumspecific detector. *Se pu*= Chinese journal of chromatography, *17*(3), 240-244.
- Mervyn, L. (1985). The Dictionary of Minerals: The Complete Guide to Minerals and Mineral Therapy. Melbourne, Australia: Lothian Publishing.
- Ministry of Agriculture (2010). Ministry of Agriculture of People's Republic of China Bulletin No. 1224–2010. The safety use standard of feed additives. Beijing, China.
- Nakagawa, T.; Aoyama, E.; Hasegawa, N.; Kobayashi, N. and Tanaka, H. (1989). High-performance liquid chromatographyfluorometric determination of selenium based on selenotrisulfide formation reaction. Analytical Chemistry, 61(3), 233-236.
- Pedersen, G.A. and Larsen, E.H. (1997). Speciation of four selenium compounds using high performance liquid chromatog-

raphy with on-line detection by inductively coupled plasma mass spectrometry or flame atomic absorption spectrometry. Fresenius' journal of analytical chemistry, *358*(5), 591-598.

- Rotruck, J.T.; Pope, A.L.; Ganther, H.E.; Swanson, A.B.; Hafeman, D.G. and Hoekstra, W.G. (1973). Selenium: biochemical role as a component of glutathione peroxidase. Science.; 179(4073): 58890. https: doi. org/10.1126 science. 179. 4073. 588PMID:4686466.
- Surai, P.F. and Dvorska, J.E. (2002). Effect of selenium and vitamin E on lipid peroxidation in thigh muscle tissue of broiler breeder hens during storage. Arch Geflugelkd.; 66: 120.
- Tanaka, H.; Nakagawa, T.; Aoyama, E.; Chikuma, M.; Nakayama, M.; Tanaka, T. and Ito, K. (1988). Determination of Selenium in Water Samples by HPLC with Fluorescent Detection. In Studies in Environmental Science (Vol. 34, pp. 347-351). Elsevier.
- Zha, L.Y.; Xu, Z.R.; Wang, M.Q. and Gu, L.Y. (2008). Chromium nanoparticle exhibits higher absorption efficiency than chromium picolinate and chromium chloride in Caco-2 cell monolayers. J. Anim. Physiol. Anim. Nutr. (Berl); 92(2): 131–40. https:// doi. org/ 10.1111 /j.1439-0396. 2007. 00718. xP MID: 18336409.
- Zhang, J.S.; Gao, X.Y.; Zhang, L.D. and Bao, Y.P. (2001). Biological effects of a nano red elemental selenium. Biofactors; 15 (1): 27–38. https: doi. org/ 10.1002 biof. 5520150103PMID:11673642.